003216589045 (EU)
0203393 8531 (UK)
(408)780-0908 (US)
info@gentaur.com
www.caslab.com

Rat SP-A1 (Pulmonary surfactant-associated protein A1) CLIA Kit

Cat #: G-EC-01751

Species: Rat

Sensitivity: 7.5pg/mL

Type: Sandwich

Range: 12.5~800pg/mL

Detection: Chemiluminescence

Duration: 3.5h

Principle:
This CLIA kit uses the Sandwich-CLIA principle. The micro CLIA plate provided in this kit has been pre-coated with an antibody specific to Rat SP-A1. Standards or samples are added to the micro CLIA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Rat SP-A1 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Rat SP-A1, biotinylated detection antibody and Avidin-HRP conjugate will appear fluorescence. The Relative light unit (RLU) value is measured by the Chemiluminescence immunoassay analyzer. The RLU value is positively associated with the concentration of Rat SP-A1. You can calculate the concentration of Rat SP-A1 in the samples by comparing the RLU value of the samples to the standard curve.

Specificity:
This kit recognizes Rat SP-A1 in samples. No significant cross-reactivity or interference between Rat SP-A1 and analogues was observed.

UniProt ID:

Gene: SP-A1

Synonyms:

Specimen: Serum, plasma and other biological fluids

Application: This CLIA kit applies to the in vitro quantitative determination of Rat SP-A1 concentrations in Serum, plasma and other biological flu

Research Areas: Signal transduction

Storage:
An unopened kit can be stored at 2-8°C for 1 month. If the kit is not supposed to be used within 1 month, store the items separately according to the storage table in the manual once the kit is received.