003216589045 (EU)
0203393 8531 (UK)
(408)780-0908 (US)
info@gentaur.com
www.caslab.com

Human MMP-12 (Matrix Metalloproteinase 12) ELISA Kit

Cat #: G-EC-02282

Species: Human

Sensitivity: 0.19ng/mL

Type: Sandwich

Range: 0.31~20ng/mL

Detection: Colormetric

Duration: 3.5h

Principle:
This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human MMP-12. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human MMP-12 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human MMP-12, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 ± 2 nm. The OD value is proportional to the concentration of Human MMP-12. You can calculate the concentration of Human MMP-12 in the samples by comparing the OD of the samples to the standard curve.

Specificity:
This kit recognizes Human MMP-12 in samples. No significant cross-reactivity or interference between Human MMP-12 and analogues was observed.

UniProt ID: P39900

Gene: MMP-12

Synonyms: MMP12, HME, ME, MME, Macrophage Metalloelastase, Macrophage Elastase

Specimen: Serum, plasma and other biological fluids

Application: This ELISA kit applies to the in vitro quantitative determination of Human MMP-12 concentrations in Serum, plasma and other biolo

Research Areas: Cancer, Cardiovascular, Signal transduction

Storage:
An unopened kit can be stored at 2-8°C for 1 month. If the kit is not supposed to be used within 1 month, store the items separately according to the storage table in the manual once the kit is received.