003216589045 (EU)
0203393 8531 (UK)
(408)780-0908 (US)
info@gentaur.com
www.caslab.com

Human BACE1 (Beta-Site APP Cleaving Enzyme 1) ELISA Kit

Cat #: G-EC-02329

Species: Human

Sensitivity: 0.38ng/mL

Type: Sandwich

Range: 0.63~40ng/mL

Detection: Colormetric

Duration: 3.5h

Principle:
This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human BACE1. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human BACE1 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human BACE1, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 ± 2 nm. The OD value is proportional to the concentration of Human BACE1. You can calculate the concentration of Human BACE1 in the samples by comparing the OD of the samples to the standard curve.

Specificity:
This kit recognizes Human BACE1 in samples. No significant cross-reactivity or interference between Human BACE1 and analogues was observed.

UniProt ID: P56817

Gene: BACE1

Synonyms: ASP2, HSPC104, Beta Secretase 1, Memapsin 2, membrane-associated aspartic protease 2, aspartyl protease 2

Specimen: Serum, plasma and other biological fluids

Application: This ELISA kit applies to the in vitro quantitative determination of Human BACE1 concentrations in Serum, plasma and other biologi

Research Areas: Cancer, Cell Biology, Neuroscience

Storage:
An unopened kit can be stored at 2-8°C for 1 month. If the kit is not supposed to be used within 1 month, store the items separately according to the storage table in the manual once the kit is received.