003216589045 (EU)
0203393 8531 (UK)
(408)780-0908 (US)
info@gentaur.com
www.caslab.com

# Human PCDH16 (Protocadherin Beta 16) ELISA Kit

**Cat #:** G-EC-02787

**Species:** Human

**Sensitivity:** 0.09ng/mL

**Type:** Sandwich

**Range:** 0.16~10ng/mL

**Detection:** Colormetric

**Duration:** 3.5h

**Principle:**
This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human PCDH16. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human PCDH16 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human PCDH16, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 ± 2 nm. The OD value is proportional to the concentration of Human PCDH16. You can calculate the concentration of Human PCDH16 in the samples by comparing the OD of the samples to the standard curve.

**Specificity:**
This kit recognizes Human PCDH16 in samples. No significant cross-reactivity or interference between Human PCDH16 and analogues was observed.

**UniProt ID:** Q9NRJ7

**Gene:** PCDH16

**Synonyms:** PCDHB16, ME1, PCDH-BETA16, PCDH3X, PCDHB8a

**Specimen:** Serum, plasma and other biological fluids

**Application:** This ELISA kit applies to the in vitro quantitative determination of Human PCDH16 concentrations in Serum, plasma and other biolog

**Research Areas:** Neuroscience, Signal transduction

**Storage:**
An unopened kit can be stored at 2-8°C for 1 month. If the kit is not supposed to be used within 1 month, store the items separately according to the storage table in the manual once the kit is received.