003216589045 (EU)
0203393 8531 (UK)
(408)780-0908 (US)
info@gentaur.com
www.caslab.com

Human PECAM1/CD31 (Platelet/Endothelial Cell Adhesion Molecule 1) ELISA Kit

Cat #: G-EC-03193

Species: Human

Sensitivity: 18.75pg/mL

Type: Sandwich

Range: 31.25~2000pg/mL

Detection: Colormetric

Duration: 3.5h

Principle:
This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human PECAM1/CD31. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human PECAM1/CD31 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human PECAM1/CD31, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 ± 2 nm. The OD value is proportional to the concentration of Human PECAM1/CD31. You can calculate the concentration of Human PECAM1/CD31 in the samples by comparing the OD of the samples to the standard curve.

Specificity:
This kit recognizes Human PECAM1/CD31 in samples. No significant cross-reactivity or interference between Human PECAM1/CD31 and analogues was observed.

UniProt ID: P16284

Gene: PECAM1/CD31

Synonyms: Platelet endothelial cell adhesion molecule, PECAM-1, cluster of differentiation 31, CD31, CD31/EndoCAM, GPIIA', PECA1, PECAM-1, e

Specimen: Serum, plasma and other biological fluids

Application: This ELISA kit applies to the in vitro quantitative determination of Human PECAM1/CD31 concentrations in Serum, plasma and other

Research Areas: Cancer, Cardiovascular, Immunology, Signal transduction, Stem cells

Storage:
An unopened kit can be stored at 2-8°C for 1 month. If the kit is not supposed to be used within 1 month, store the items separately according to the storage table in the manual once the kit is received.